

Trinseo Medical Grade Plastics Biocompatibility Testing Data Sheet

All of Trinseo's Commercial Medical Grade products have been evaluated for their ability to successfully pass a standard battery of biocompatibility testing. These evaluations have been conducted by the contract labs of North American Science Associates Incorporated, Northwood, Ohio (NAMSA) based on the guidelines of the International Organization of Standardization (ISO 10993). In addition, these biocompatibility studies were conducted in accordance with the provisions of the FDA Good Laboratory Practice (GLP) Regulations (21 CFR, Part 58). The table below lists the tests, with applicable test method, which Trinseo products must pass in order to be sold by Trinseo as a medical grade product.

TEST	METHOD ^a
In Vitro Hemolysis: Extraction Method	ISO 10993 Part 4/Modified ASTM
In Vitro Hemolysis: Direct Contact Method	ISO 10993 Part 4/Modified ASTM
Cytotoxicity, Elution Method	ISO 10993 Part 5
Muscle Implantation – 1 week	ISO 10993 Part 6
Guinea Pig Maximization Sensitization	ISO 10993 Part 10
Intracutaneous	ISO 10993 Part 10
Systemic Toxicity	ISO 10993 Part 11
Physicochemical Test	USP<661>

^a ISO = International Organization of Standardization 10993: Biological Evaluation of Medical Devices

^a ASTM = American Society of Testing Materials

^a USP = United States Pharmacopeia

These studies are performed on pellet samples of medical grade resin. The purpose of this testing is to ascertain whether the Trinseo materials cause any health effects. This testing is intended only as a preliminary qualification step and does not take into account processes or conditions of use by the customer; instead, the medical device manufacturer is responsible for evaluating its product based upon such factors as manufacturing processes and conditions of use. Trinseo LLC does not submit any finished devices or articles for testing.

Trinseo can provide detailed information with regard to biocompatibility testing and regulatory compliance for its products. New products may require additional time for the biocompatibility evaluation. Biocompatibility and compliance letters on new products will be provided upon request after final product testing and approval is completed.

The principles of Responsible Care® and sustainability influence the production of printed literature for Trinseo S.A. and its affiliated companies. As a contribution towards the protection of our environments, Trinseo's printed literature is produced in small quantities and on paper containing recovered/post-consumer fiber and using 100 percent soy-based ink whenever possible.

Product Stewardship

Trinseo and its affiliated companies have a fundamental concern for all who make, distribute, and use their products and for the environment in which we live. This concern is the basis for our Product Stewardship philosophy by which we assess the safety, health, and environmental information on our products so that appropriate steps may be taken to protect employee and public health and our environment. The success of our product stewardship program rests with each and every individual involved with Trinseo products – from the initial concept and research, to manufacture, use, sale, disposal, and recycle of each product.

Customer Notice

Customers are responsible for reviewing their manufacturing processes and their applications of Trinseo products from the standpoint of human health and environmental quality to ensure that Trinseo products are not used in ways for which they are not suitable. Trinseo personnel are available to answer questions and to provide reasonable technical support. Trinseo product literature, including safety data sheets, should be consulted prior to the use of Trinseo products. Current safety data sheets are available from Trinseo.

No freedom from infringement of any patent owned by Trinseo or others is to be inferred. Because use conditions and applicable laws may differ from one location to another and may change with time, the customer is responsible for determining whether products and the information in this document are appropriate for the customer's use and for ensuring that the customer's workplace and disposal practices are in compliance with applicable legal requirements. Although the information herein is provided in good faith and was believed to be accurate when prepared, Trinseo assumes no obligation or liability for the information in this document.

NOTICE REGARDING MEDICAL APPLICATION RESTRICTIONS

TRINSEO REQUESTS THAT CUSTOMERS REFER TO TRINSEO'S MEDICAL APPLICATION POLICY [HTTP://WWW.TRINSEO.COM/MEDICAL.HTM](http://www.trinseo.com/MEDICAL.HTM) BEFORE CONSIDERING THE USE OF TRINSEO PRODUCTS IN MEDICAL APPLICATIONS. THE RESTRICTIONS AND DISCLAIMERS SET FORTH IN THAT POLICY ARE INCORPORATED BY REFERENCE.

For more information on products, innovations, expertise, and other services available from Trinseo, visit www.trinseo.com, or in the U.S. contact us at +1-855-TRINSEO (+1-855-874-6736).

DISCLAIMER

TRINSEO MAKES NO WARRANTIES, EITHER EXPRESS OR IMPLIED, IN THIS DOCUMENT; ALL IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE (INCLUDING MEDICAL APPLICATIONS) ARE EXPRESSLY EXCLUDED. SINCE THE CONDITIONS AND METHODS OF USE OF THE INFORMATION AND PRODUCTS REFERRED TO ARE BEYOND TRINSEO'S KNOWLEDGE AND CONTROL, TRINSEO DISCLAIMS ANY AND ALL LIABILITY FOR LOSSES OR DAMAGES THAT MAY RESULT FROM RELIANCE ON THE INFORMATION OR USE OF THE PRODUCTS DESCRIBED HEREIN. TRINSEO MAKES NO WARRANTIES, EXPRESS OR IMPLIED, THAT THE USE OF ANY TRINSEO PRODUCT WILL BE FREE FROM ANY INFRINGEMENT CLAIMS.

GENERAL NOTICE

Any photographs of end-use applications in this document represent potential end-use applications but do not necessarily represent current commercial applications, nor do they represent an endorsement by Trinseo of the actual products. Further, these photographs are for illustration purposes only and do not reflect either an endorsement or sponsorship of any other manufacturer for a specific potential end-use product or application, or for Trinseo, or for specific products manufactured by Trinseo. If products are described as "experimental" or "developmental": (1) product specifications may not be fully determined; (2) analysis of hazards and caution in handling and use are required; (3) there is greater potential for Trinseo to change specifications and/or discontinue production, and (4) although Trinseo may from time to time provide samples of such products, Trinseo is not obligated to supply or otherwise commercialize such products for any use or application whatsoever.

For additional information not covered by the content of this document or to ensure you have the latest version of this document available, please refer to our website at www.trinseo.com/contact.

Copyright© Trinseo (2019) All rights reserved.

™Trademark of Trinseo S.A. or its affiliates

®Responsible Care is a service mark of the American Chemistry Council

Follow us at:

